

| | Confidence interval of positive coincidence rate | Confidence interval of negative coincidence rate | Confidence interval of total coincidence rate |
|---|--|--|---|
| Test reagent results and clinical diagnosis results | 91.54% | 97.02% | 95.09% |
| Test reagent results and nucleic acid test results | 89.55% | 82.80% | 84.39% |

In summary, this clinical study showed that 2019 novel coronavirus (2019-nCoV) IgG/IgM antibody test kit (colloidal gold method) researched and developed by Vemur Medical Technology Co., Ltd. is of great value in clinical application for