

Laboratory testing for coronavirus disease (COVID-19) in suspected human cases.

Interim guidance
19 March 2020



Background

This document provides interim guidance to laboratories and stakeholders involved in COVID-19 virus laboratory testing of patients.

It is based in part on the interim guidance on laboratory testing for Middle East Respiratory Syndrome (MERS) coronavirus.¹⁻⁶ Information on human infection with the COVID-19 virus is evolving and WHO continues to monitor developments and revise recommendations as necessary. This document will be revised as new information becomes available. Feedback is welcome and can be sent to WHElab@who.int.

The virus has now been named SARS-CoV-2 by the International Committee of Taxonomy of Viruses (ICTV)⁷ (2). This virus can cause the disease named coronavirus disease 2019 (COVID-19). WHO refers to the virus as COVID-19 virus in its current documentation.

Laboratory testing guiding principles for patients who meet the suspect case definition.

The decision to test should be based on clinical and epidemiological factors and linked to an assessment of the likelihood of infection. PCR testing of asymptomatic or mildly symptomatic contacts can be considered in the assessment of individuals who have had contact with a COVID-19 case. Screening protocols should be adapted to the local situation. The case definitions are being regularly reviewed and updated as new information becomes available. For the WHO suspected case definition see: Global Surveillance for human infection with coronavirus disease (COVID-2019).⁸

Rapid collection and testing of appropriate specimens from patients meeting the suspected case definition for COVID-19 is a priority for clinical management and outbreak control and should be guided by a laboratory expert. Suspected cases should be screened for the virus with nucleic acid amplification tests (NAAT), such as RT-PCR.

If testing for COVID-19 is not yet available nationally, specimens should be referred. A list of WHO reference laboratories providing confirmatory testing for COVID-19 and shipment instructions are [available](#).

If case management requires, patients should be tested for other respiratory pathogens using routine laboratory procedures, as recommended in local management guidelines for community-acquired pneumonia. Additional testing should not delay testing for COVID-19. As co-infections can occur, all patients that meet the suspected case definition should be tested for COVID-19 virus regardless of whether another respiratory pathogen is found.

In an early study in Wuhan, the mean incubation period for COVID-19 was 5.2 days among 425 cases, though it varies widely between individuals.⁹⁻¹¹ Virus shedding patterns are not yet well understood and further investigations are needed to better understand the timing, compartmentalization, and quantity of viral shedding to inform optimal specimen collection. Although respiratory samples have the greatest yield, the virus can be detected in other specimens, including stool and blood.^{12,14} Local guidelines on informed consent should be followed for specimen collection, testing, and potentially future research.

Specimen collection and shipment

Safety procedures during specimen collection

Ensure that adequate standard operating procedures (SOPs) are in use and that staff are trained for appropriate specimen collection, storage, packaging, and transport. All specimens collected for laboratory investigations should be regarded as potentially infectious.

Ensure that health care workers who collect specimens adhere rigorously to infection prevention and control guidelines. Specific WHO interim guidance has been published.¹⁶

Box 1. Biosafety practices in the laboratory

Testing on clinical specimens from patients meeting the suspected case definition should be performed in appropriately equipped laboratories by staff trained in the relevant technical and safety procedures. National guidelines on laboratory biosafety should be followed in all circumstances. There is still limited information on the risk posed by COVID-19, but all procedures should be undertaken based on a risk assessment. Specimen handling for molecular testing would require BSL-2 or equivalent facilities. Attempts to culture the virus require BSL-3 facilities at minimum.

For more information related to COVID-19 risk assessment, see: [WHO interim guidance for laboratory biosafety related to 2019-nCoV](#). Samples that are potentially infectious materials (PIM) for polio need to be handled and stored as described in WHO document [Guidance to minimize risks for facilities collecting, handling or storing materials potentially infectious for polioviruses \(PIM Guidance\)](#). For general laboratory biosafety guidelines, see the [WHO Laboratory Biosafety Manual, 3rd edition](#) before the 4th edition is released.

Specimens to be collected

At minimum, respiratory material should be collected:

- upper respiratory specimens: nasopharyngeal and oropharyngeal swab or wash in ambulatory patients
- and/or lower respiratory specimens: sputum (if produced) and/or endotracheal aspirate or bronchoalveolar lavage in patients with more severe respiratory disease. (Note high risk of aerosolization; adhere strictly to infection prevention and control procedures).

Additional clinical specimens may be collected as COVID-19 virus has been detected in blood and stool, as had the coronaviruses responsible for SARS and MERS.^{12,14,17-19} The duration and frequency of shedding of COVID-19 virus in stool and potentially in urine is unknown. In case of patients who are deceased, consider autopsy material including lung tissue. In surviving patients, paired serum (acute and convalescent) can be useful to retrospectively define cases as serological assays become available.

Further recommendations on materials to collect, including the testing of asymptomatic individuals, can be found in Table 1.

Packaging and shipment of clinical specimens

Specimens for virus detection should reach the laboratory as soon as possible after collection. Correct handling of specimens during transportation is essential. Specimens that can be delivered promptly to the laboratory can be stored and shipped at 2-8°C. When there is likely to be a delay in specimens reaching the laboratory, the use of viral transport medium is strongly recommended. Specimens may be frozen to -20°C or ideally -70°C and shipped on dry ice if further delays are expected (see Table 2). It is important to avoid repeated freezing and thawing of specimens.

Transport of specimens within national borders should comply with applicable national regulations. International transport of potentially COVID-19 virus containing samples should follow the UN Model Regulations, and any other applicable regulations depending on the mode of transport being used. More information may be found in the WHO Guidance on regulations for the Transport of Infectious Substances 2019-2020²² and WHO interim guidance for laboratory biosafety related to coronavirus disease.¹⁶

Ensure good communication with the laboratory and provide needed information.

Alerting the laboratory before sending specimens encourages proper and timely processing of samples and timely reporting. Specimens should be correctly labelled and accompanied by a diagnostic request form (template provided in Annex I).

Laboratory testing for COVID-19 virus

Laboratories undertaking testing for COVID-19 virus should adhere strictly to appropriate biosafety practices.

Nucleic acid amplification tests (NAAT) for COVID-19 virus.

Routine confirmation of cases of COVID-19 is based on detection of unique sequences of virus RNA by NAAT such as real-time reverse-transcription polymerase chain reaction (rRT-PCR) with confirmation by nucleic acid sequencing when necessary. The viral genes targeted so far include the N, E, S and RdRP genes. Examples of protocols used may be found here. RNA extraction should be done in a biosafety cabinet in a BSL-2 or equivalent facility. Heat treatment of samples before RNA extraction is not recommended.

Laboratory confirmation of cases by NAAT in areas with no known COVID-19 virus circulation.

To consider a case as laboratory-confirmed by NAAT in an area with no COVID-19 virus circulation, one of the following conditions need to be met:

- A positive NAAT result for at least two different targets on the COVID-19 virus genome, of which at least one target is preferably specific for COVID-19 virus using a validated assay (as at present no other SARS-like coronaviruses are circulating in the human population it can be debated whether it must be COVID-19 or SARS-like coronavirus specific); OR
- One positive NAAT result for the presence of betacoronavirus, and COVID-19 virus further identified by sequencing partial or whole genome of the virus as long as the sequence target is larger or different from the amplicon probed in the NAAT assay used.

When there are discordant results, the patient should be resampled and, if appropriate, sequencing of the virus from the original specimen or of an amplicon generated from an appropriate NAAT assay, different from the NAAT assay initially used, should be obtained to provide a reliable test result. Laboratories are urged to seek confirmation of any surprising results in an international reference laboratory.

Laboratory-confirmed case by NAAT in areas with established COVID-19 virus circulation.

In areas where COVID-19 virus is widely spread a simpler algorithm might be adopted in which, for example, screening by rRT-PCR of a single discriminatory target is considered sufficient.

One or more negative results do not rule out the possibility of COVID-19 virus infection. A number of factors could lead to a negative result in an infected individual, including:

- poor quality of the specimen, containing little patient material (as a control, consider determining whether there is adequate human DNA in the sample by including a human target in the PCR testing).
- the specimen was collected late or very early in the infection.
- the specimen was not handled and shipped appropriately.

- technical reasons inherent in the test, e.g. virus mutation or PCR inhibition.

If a negative result is obtained from a patient with a high index of suspicion for COVID-19 virus infection, particularly when only upper respiratory tract specimens were collected, additional specimens, including from the lower respiratory tract if possible, should be collected and tested.

Each NAAT run should include both external and internal controls, and laboratories are encouraged to participate in external quality assessment schemes when they become available. It is also recommended to laboratories that order their own primers and probes to perform entry testing/validation on functionality and potential contaminants.

Serological testing

Serological surveys can aid investigation of an ongoing outbreak and retrospective assessment of the attack rate or extent of an outbreak. In cases where NAAT assays are negative and there is a strong epidemiological link to COVID-19 infection, paired serum samples (in the acute and convalescent phase) could support diagnosis once validated serology tests are available. Serum samples can be stored for these purposes.

Cross reactivity to other coronaviruses can be challenging,²² but commercial and non-commercial serological tests are currently under development. Some studies with COVID-19 serological data on clinical samples have been published.^{23,24}

Viral sequencing

In addition to providing confirmation of the presence of the virus, regular sequencing of a percentage of specimens from clinical cases can be useful to monitor for viral genome mutations that might affect the performance of medical countermeasures, including diagnostic tests. Virus whole genome sequencing can also inform molecular epidemiology studies. Many public-access databases for deposition of genetic sequence data are available, including GISAID, which is intended to protect the rights of the submitting party.²⁵

Viral culture

Virus isolation is not recommended as a routine diagnostic procedure.

Reporting of cases and test results

Laboratories should follow national reporting requirements. In general, all test results, positive or negative, should be immediately reported to national authorities. States Parties to the IHR are reminded of their obligations to share with WHO relevant public health information for events for which they notified WHO, using the decision instrument in Annex 1 of the IHR (2005).²⁶

Research toward improved detection of COVID-19 virus.

Many aspects of the virus and disease are still not understood. A better understanding will be needed to provide improved guidance. For example:

Viral dynamics: optimal timing and type of clinical material to sample for molecular testing-

- Dynamic of immunological response
- Disease severity in various populations, e.g. by age.
- The relationship between viral concentration and disease severity.
- The duration of shedding, and relation to clinical picture (e.g. clinical recovery occurs with viral clearing, or shedding persists despite clinical improvement).
- Development and validation of useful serological assays.
- Comparative studies of available molecular and serological assays.
- Optimal percentage of positive cases that requires sequencing to monitor mutations that might affect the performance of molecular tests.
- WHO encourages the sharing of data to better understand and thus manage the COVID-19 outbreak, and to develop countermeasures.

Table 1. Specimens to be collected from symptomatic patients and contacts

	Test	Type of sample	Timing
Patient	NAAT	<p>Lower respiratory tract</p> <ul style="list-style-type: none"> - sputum - aspirate - lavage <p>Upper respiratory tract</p> <ul style="list-style-type: none"> - nasopharyngeal and oropharyngeal swabs - nasopharyngeal wash/nasopharyngeal aspirate. <p>Consider stools, whole blood, urine, and if diseased, material from autopsy.</p>	Collect on presentation. Possibly repeated sampling to monitor clearance. Further research needed to determine effectiveness and reliability of repeated sampling.
Patient	Serology	Serum for serological testing once validated and available.	Paired samples are necessary for confirmation with the initial sample collected in the first week of illness and the second ideally collected 2-4 weeks later (optimal timing for convalescent sample needs to be established).
Contact in health-care centre associated outbreaks or other settings where contacts have symptoms, or where asymptomatic contacts have had high-intensity contact with a COVID-19 case.	NAAT	Nasopharyngeal and oropharyngeal swabs.	Within incubation period of last documented contact.
	Serology	Serum for serological testing once validated and available.	Baseline serum taken as early as possible within incubation period of contact and convalescent serum taken 2-4 weeks after last contact (optimal timing for convalescent sample needs to be established).

Table 2. Specimen collection and storage (adapted from^{4, 27, 28})

Specimen type	Collection materials	Storage temperature until testing in-country laboratory	Recommended temperature for shipment according to expected shipment time
Nasopharyngeal and oropharyngeal swab	Dacron or polyester flocked swabs*	2-8 °C	2-8 °C if ≤5 days -70 °C (dry ice) if >5 days
Bronchoalveolar lavage	Sterile container *	2-8 °C	2-8 °C if ≤2 days -70 °C (dry ice) if >2 days
(Endo)tracheal aspirate, nasopharyngeal or nasal wash/aspirate	Sterile container *	2-8 °C	2-8 °C if ≤2 days -70 °C (dry ice) if >2 days
Sputum	Sterile container	2-8 °C	2-8 °C if ≤2 days -70 °C (dry ice) if >2 days
Tissue from biopsy or autopsy including from lung.	Sterile container with saline or VTM.	2-8 °C	2-8 °C if ≤24 hours -70 °C (dry ice) if >24 hours
Serum	Serum separator tubes (adults: collect 3-5 ml whole blood).	2-8 °C	2-8 °C if ≤5 days -70 °C (dry ice) if >5 days
Whole blood	Collection tube	2-8 °C	2-8 °C if ≤5 days -70 °C (dry ice) if >5 days
Stool	Stool container	2-8 °C	2-8 °C if ≤5 days -70 °C (dry ice) if >5 days
Urine	Urine collection container	2-8 °C	2-8 °C if ≤5 days -70 °C (dry ice) if >5 days

* For transport of samples for viral detection, use viral transport medium (VTM) containing antifungal and antibiotic supplements. Avoid repeated freezing and thawing of specimens. If VTM is not available sterile saline may be used instead (in which case, duration of sample storage at 2-8 °C may be different from what is indicated above).

Aside from specific collection materials indicated in the table also assure other materials and equipment are available: e.g. transport containers and specimen collection bags and packaging, coolers, and cold packs or dry ice, sterile blood-drawing equipment (e.g. needles, syringes and tubes), labels and permanent markers, PPE, materials for decontamination of surfaces, etc.

References

1. Laboratory testing for Middle East Respiratory Syndrome coronavirus, interim guidance (revised), January 2019, WHO/MERS/LAB/15.1/Rev1/2019, World Health Organization, 2018. (<https://apps.who.int/iris/bitstream/handle/10665/259952/WHO-MERS-LAB-15.1-Rev1-2018-eng.pdf;jsessionid=08939A780A5A4552EC8279D0E6D650E4?sequence=1>)
2. Managing epidemics, key facts about major deadly diseases. Geneva: World Health Organization; 2018. (<https://apps.who.int/iris/handle/10665/272442>).
3. WHO Global Influenza Surveillance Network Manual for the laboratory diagnosis and virological surveillance of influenza, WHO, 2011 (https://www.who.int/influenza/gisrs_laboratory/manual_diagnosis_surveillance_influenza/en/).
4. Protocol to investigate non-seasonal influenza and other emerging acute respiratory diseases. Geneva: World Health Organization; 2018. (https://www.who.int/influenza/resources/publications/outbreak_investigation_protocol/en/).
5. WHO Recommended Surveillance Standards WHO/CDS/CSR/ISR/99.2 (<https://www.who.int/csr/resources/publications/surveillance/whocdscr992.pdf>).
6. Guideline for the collection of clinical specimens during field investigation of outbreaks WHO/CDS/CSR/EDC/200.4 (https://www.who.int/ihr/publications/WHO_CDS_CSR_EDC_2000_4/en/).
7. Severe acute respiratory syndrome-related coronavirus: The species and its viruses – a statement of the Coronavirus Study Group. Alexander E. Gorbalenya, Susan C. Baker, Ralph S. Baric, Raoul J. de Groot, Christian Drosten, Anastasia A. Gulyaeva, Bart L. Haagmans, Chris Lauber, Andrey M. Leontovich, Benjamin W. Neuman, Dmitry Penzar, Stanley Perlman, Leo L.M. Poon, Dmitry Samborskiy, Igor A. Sidorov, Isabel Sola, John Ziebuhr. Severe acute respiratory syndrome-related coronavirus: The species and its viruses – a statement of the Coronavirus Study Group. bioRxiv 2020.02.07.937862; doi: <https://doi.org/10.1101/2020.02.07.937862>

8. WHO laboratory biosafety manual, third edition. Geneva: World Health Organization; 2004. (http://www.who.int/csr/resources/publications/bio_safety/WHO_CDS_CSR_LYO_2004_11/en/).
9. Global Surveillance for human infection with coronavirus disease (COVID-2019), Interim guidance, Geneva, World Health Organization, 2020. ([https://www.who.int/publications-detail/global-surveillance-for-human-infection-with-novel-coronavirus-\(2019-ncov\)](https://www.who.int/publications-detail/global-surveillance-for-human-infection-with-novel-coronavirus-(2019-ncov))).
10. Qun Li, Xuhua Guan, Peng Wu, et al. Early Transmission Dynamics in Wuhan, China, of Novel Coronavirus–Infected Pneumonia. *N Eng J Med*. January 29, 2020.
11. Wei-jie Guan, Zheng-yi Ni, Yu Hu, Wen hua Liang, Chun-quan Ou, Jian xing He, et al. Clinical characteristics of 2019 novel coronavirus infection in China. medRxiv preprint doi: <https://doi.org/10.1101/2020.02.06.20020974>.
12. Nathalie M Linton, Tetsuro Kobayashi, Yichi Yang, Katsuma Hayashi, Andrei R. Akhmetzhanov, Sung mok Jung, et al. Incubation Period and Other Epidemiological Characteristics of 2019 Novel Coronavirus Infections with Right Truncation: A Statistical Analysis of Publicly Available Case Data. *J. Clin. Med*. 2020, 9(2), 538; <https://doi.org/10.3390/jcm9020538>.
13. Xu Kaijin, Cai Hongliu, Shen Yihong, et al. Management of Corona Virus Disease 19 (COVID-19): the Zhejiang Experience. *Jour Zhejiang Univ*. DOI: 10.3785/j.issn.1008-9292.2020.02.02.
14. Wei Zhang, Rong-Hui Du, Bei Li, Xiao Shuang Zheng, Xing-Lou Yang, Ben Hu, et al. Molecular and serological investigation of 2019-nCoV infected patients: implication of multiple shedding routes, *Emerging Microbes & Infections* 2020 9:1, 386-389 (<https://www.tandfonline.com/doi/full/10.1080/22221751.2020.1729071>).
15. Infection prevention and control during health care when novel coronavirus (nCoV) infection is suspected, interim guidance, January 2020. Geneva: World Health Organization; 2020. ([https://www.who.int/publications-detail/infection-prevention-and-control-during-health-care-when-novel-coronavirus-\(ncov\)-infection-is-suspected-20200125](https://www.who.int/publications-detail/infection-prevention-and-control-during-health-care-when-novel-coronavirus-(ncov)-infection-is-suspected-20200125)).
16. Laboratory biosafety guidance related to the novel coronavirus (2019-nCoV), World Health Organization; 2020 (https://www.who.int/docs/default-source/coronaviruse/laboratory-biosafety-novel-coronavirus-version-1-1.pdf?sfvrsn=912a9847_2).
17. Shi X, Gong E, Gao D, et al. Severe acute respiratory syndrome associated coronavirus is detected in intestinal tissues of fatal cases. *Am J Gastroenterol*. 2005;100 (1):169–176.
18. Zhou J, Li C, Zhao G, et al. Human intestinal tract serves as an alternative infection route for Middle East respiratory syndrome coronavirus. *Sci Adv*. 2017;3(11).
19. Ding Y, He L, Zhang Q, et al. Organ distribution of severe acute respiratory syndrome (SARS) associated coronavirus (SARS-CoV) in SARS patients: implications for pathogenesis and virus transmission pathways. *J. Pathol*. 2004;203(2):622–630.
20. Guidance on regulations for the transport of infectious substances 2019–2020. Geneva: World Health Organization; 2019. (<https://www.who.int/ihr/publications/WHO-WHE-CPI-2019.20/en/>).
21. Guidance to minimize risks for facilities collecting, handling or storing materials potentially infectious for polioviruses (PIM Guidance). Geneva: World Health Organization; 2018. <http://polioeradication.org/wp-content/uploads/2016/07/PIM-guidance-20190122-EN.pdf>.
22. Meyer B, Drosten C, Müller MA. Serological assays for emerging coronaviruses: challenges and pitfalls. *Virus Res*. 2014 Dec 19;194:175-83.
23. Bai Shaoli, Wang Jianyun, Zhou (Yingquan) Yu Desheng, Gao Xiaomin, Li Lingling, Yang Fan. Analysis of the first family epidemic situation of new coronavirus pneumonia in Gansu Province. *Chinese Journal of Preventive medicine*, 2020, 54.
24. Shu-Yuan Xiao, Yingjie Wu, Juan Li, Evolving status of the 2019 novel coronavirus infections: proposal of conventional serologic assays for disease diagnostics and infection monitoring. 2020, *J Med Virol*. 2020;1-4.
25. GISAID.org (<https://www.gisaid.org/>), accessed on 19 February 2020
26. World Health Organization. International Health Regulations (2005), third edition. Geneva: World Health Organization; 2016 (<http://www.who.int/ihr/publications/9789241580496/en/>).
27. Protocol to investigate non-seasonal influenza and other emerging acute respiratory diseases. Geneva: World Health Organization; 2018. (https://www.who.int/influenza/resources/publications/outbreak_investigation_protocol/en/).
28. Bruce et al. *JCM*. 2011. Evaluation of Swabs, Transport Media, and Specimen Transport Conditions for Optimal Detection of Viruses by PCR.

Acknowledgements

The following people contributed to the drafting of the evolving versions of this guidance document: Katrin Leitmeyer, European Center for Disease Control, Maria Zambon, Public Health England, UK; Christian Drosten, Charité - Universitätsmedizin Berlin, Germany; Marion Koopmans, Erasmus MC, Rotterdam, The Netherlands; Leo Poon, Hong Kong University, China, Hong Kong SAR; George Gao, Chinese CDC, China. WHO: Karen Nahapetyan, Francis Inbanathan, Dmitriy Pereyaslov, Christine Uhlenhaut, Varja Grabovac, Katelijin Vandemaale, Magdi Samaan, Christian Fuster,

Wenqing Zhang, Lisa Stevens, Chris Oxenford, Sebastian Cognat, Kazunobu Kojima, Carmen Dolea, Caroline Brown,

Céline Barnadas, Maria Van Kerkhove, Lisa Carter, Mark D Perkins and Karin von Eije.

WHO continues to monitor the situation closely for any changes that may affect this interim guidance. Should any factors change, WHO will issue a further update. Otherwise, this interim guidance document will expire 2 years after the date of publication

Annex I

COVID-19 VIRUS LABORATORY TEST REQUEST FORM¹

Submitter information			
NAME OF SUBMITTING HOSPITAL, LABORATORY, or OTHER FACILITY*			
Physician			
Address			
Phone number			
Case definition: ²	<input type="checkbox"/> Suspected case <input type="checkbox"/> Probable case		
Patient info			
First name		Last name	
Patient ID number		Date of Birth	Age:
Address		Sex	<input type="checkbox"/> Male <input type="checkbox"/> Female <input type="checkbox"/> Unknown
Phone number			
Specimen information			
Type	<input type="checkbox"/> Nasopharyngeal and oropharyngeal swab <input type="checkbox"/> Bronchoalveolar lavage <input type="checkbox"/> Endotracheal aspirate <input type="checkbox"/> Nasopharyngeal aspirate <input type="checkbox"/> Nasal wash <input type="checkbox"/> Sputum <input type="checkbox"/> Lung tissue <input type="checkbox"/> Serum <input type="checkbox"/> Whole blood <input type="checkbox"/> Urine <input type="checkbox"/> Stool <input type="checkbox"/> Other:		
All specimens collected should be regarded as potentially infectious and you <u>must contact</u> the reference laboratory <u>before</u> sending samples. All samples must be sent in accordance with category B transport requirements.			
Please tick the box if your clinical sample is post mortem <input type="checkbox"/>			
Date of collection		Time of collection	
Priority status			
Clinical details			
Date of symptom onset:			
Has the patient had a recent history of travelling to an affected area?	<input type="checkbox"/> Yes <input type="checkbox"/> No	Country	
		Return date	
Has the patient had contact with a confirmed case?		<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> Unknown <input type="checkbox"/> Other exposure:	
Additional Comments			

© World Health Organization 2020. Some rights reserved. This work is available under the [CC BY-NC-SA 3.0 IGO](#) licence.

WHO reference number: [WHO/COVID-19/laboratory/2020.5](#)

¹ Form in accordance with ISO 15189:2012 requirements

² World Health Organization. [Global Surveillance for human infection with coronavirus disease \(COVID-19\)](#)